



Laboratory Quality Assurance Program
2700 Prosperity Avenue, Suite 250
Fairfax, VA 22031
(703) 849-8888 (703) 207-8558 fax

11/15/2007

Glenn Nawrocki
Wes Har Asbestos Analytical & Consulting Ltd.
170-2188 No. 5 Rd
Richmond, BC V6X 2T1
Canada

Lab ID# 149340

Industrial Hygiene Proficiency Analytical Testing Results

This document contains three sub-reports relating to IHPAT Round 171. The first report contains your laboratory's results listed per contaminant, per sample. The second report contains your current and 2 previous test round performance respectively (where applicable), and the final report contains summary results for all laboratories for IHPAT round 171.

Testing Results for IHPAT Round 171

This part of the report contains your laboratory's results listed per contaminant, per sample.

| Contaminant | Units | # | Result | Ref. Value | Lower Limit | Upper Limit | z-Score | Rating |
|-------------------------|-------|---|--------|------------|-------------|-------------|---------|--------|
| Asbestos / Fibers (ASB) | f/mm2 | 1 | 55 | 85 | 42 | 144 | -2.0 | A |
| | f/mm2 | 2 | 252 | 289 | 142 | 488 | -0.7 | A |
| | f/mm2 | 3 | 101 | 106 | 52 | 180 | -0.2 | A |
| | f/mm2 | 4 | 76 | 64 | 31 | 108 | 0.9 | A |
| | | | | | | | | |

Please note:

Reference value is the mean of the reference laboratories

*Lower limit = reference value - 3 standard deviations and Upper limit = reference value +3 standard deviations

*Z-score = (reported result - reference value)/standard deviation

*Asbestos is the exception because data are positively skewed therefore transformations are used to obtain approximately normal distributions.

A: Acceptable Analysis; U: Unacceptable Analysis

The acceptability of reported results is based on upper and lower performance limits. This is why a reported result may appear unacceptable according to z-score, but be identified as acceptable.

Overall Performance Summary Concluding with 171

The following table contains your laboratory's current and 2 previous test rounds performance respectively (where applicable). For more information in regard to the determination of proficiency, please see Policy Module 6B, Section 6B.2 for IHPAT and Policy Module 6C Section 6C.2 for ELPAT Lead-in-Air located at: <http://www.aiha.org/Content/LQAP/documents/accredpolicymods.htm>

| Sample | Round | Round Performance | Round Score | Proficiency Status -Three Round Score |
|----------|-------|-------------------|-------------|---------------------------------------|
| Asbestos | 169 | 4/4 | Pass | |
| | 170 | 4/4 | Pass | |
| | 171 | 4/4 | Pass | P |
| | | | | |

Please note:

The denominators represent the total number of samples analyzed.

The numerators represent the number of acceptable results.

Pass: Round Score \geq 75% Fail: Round Score $<$ 75%

P – Proficient; NP – Non-proficient.

A laboratory is rated proficient (P) for the associated FoT/Method(s), if the laboratory has a passing score for the applicable PT analyte class in two (2) of the last three (3) consecutive PT rounds. A laboratory is rated non-proficient (NP) for the applicable FoT/Method if the laboratory has failing scores for the associated PT analyte class in two (2) of the last three (3) consecutive PT rounds.

If a laboratory receives samples and does not report the data, the results will be treated as outliers.

| | | | | | | | | |
|-------------------------|---|-----|----|------|-----|-----|----|-----|
| Asbestos / Fibers (ASB) | 1 | 85 | 17 | 20.0 | 732 | 589 | 58 | 85 |
| | 2 | 289 | 58 | 20.0 | 732 | 608 | 94 | 30 |
| | 3 | 106 | 21 | 20.0 | 732 | 553 | 40 | 139 |
| | 4 | 64 | 13 | 20.0 | 732 | 670 | 30 | 32 |